



Certificate of Analysis

Description: DPD # 1 Instrument Grade Tablets
LaMotte Code Number: 6903A-M
Lot Number: 0099B
Manufacture Date: 04/2020
Expiration Date: 05/2023

<u>Test</u>	<u>Results</u>	<u>Specifications</u>
Turbidity	pass	99 – 101 % T
pH	pass	6.2 - 6.4
Actual Test :		
0.0 ppm chlorine	pass	0.0 ppm
0.5 ppm chlorine	pass	0.4 - 0.6 ppm
1.0 ppm chlorine	pass	0.9 - 1.1 ppm
3.0 ppm chlorine	pass	2.9 - 3.1 ppm
Chloramine Breakthrough	pass	< 0.1 ppm
Appearance	pass	chalk white tablets

These tablets meet established specifications.

A handwritten signature in black ink, appearing to read "Susan B. Schauber".

Susan B. Schauber
QC Manager
9/08/2020

Helping people solve analytical challenges since 1919.